




<b>LOT</b>		<b>MX 416 L</b>	<b>CONTROL L</b>							 <b>(Exp.)</b>	<b>2019-06-05</b> (YYYY-MM-DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	1.9	1.9								± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.37	2.39								± 0.15
HB HGB	g/dl	5.9	6.0								± 0.4
	g/l	59	60								± 4
	mmol/l	3.66	3.73								± 0.25
HT HCT	%	16.6	16.7								± 2.0
	l/l	0.166	0.167								± 0.020
VGM MCV	µm <sup>3</sup> ; fl	70	70								± 4
TGMH MCH	pg	24.9	25.1								± 2.0
	fmol	1.55	1.56								± 0.12
CCMH MCHC	g/dl	35.6	35.9								± 3.0
	g/l	356	359								± 30
	mmol/l	22.08	22.27								± 1.86
IDR RDW	%	N/A	14.5								± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	87								± 20
VPM MPV	µm <sup>3</sup> ; fl	N/A	9.2								± 2.0
LYMPHO	#	N/A	1.05								± 0.40
	%	N/A	55.5								± 8.0
MONO	#	N/A	0.23								± 0.20
	%	N/A	12.0								± 6.0
GRANULO	#	N/A	0.62								± 0.40
	%	N/A	32.5								± 7.0

<b>LOT</b>		<b>MX 416 N</b>	<b>CONTROL N</b>							 (Exp.)	<b>2019-06-05</b> (YYYY-MM-DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS								TOLERANCES TOLERANCE	
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.5	7.6							± 0.8	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.64	4.71							± 0.18	
HB HGB	g/dl	13.3	13.4							± 0.5	
	g/l	133	134							± 5	
	mmol/l	8.26	8.32							± 0.31	
HT HCT	%	36.7	37.2							± 2.5	
	l/l	0.367	0.372							± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	79	79							± 4	
TGMH MCH	pg	28.7	28.5							± 2.0	
	fmol	1.78	1.77							± 0.12	
CCMH MCHC	g/dl	36.3	36.0							± 3.0	
	g/l	363	360							± 30	
	mmol/l	22.53	22.36							± 1.86	
IDR RDW	%	N/A	14.0							± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	250	268							± 40	
VPM MPV	µm <sup>3</sup> ; fl	N/A	9.0							± 2.0	
LYMPHO	#	N/A	2.17							± 0.40	
	%	N/A	28.5							± 5.0	
MONO	#	N/A	0.68							± 0.30	
	%	N/A	9.0							± 4.0	
GRANULO	#	N/A	4.75							± 0.60	
	%	N/A	62.5							± 6.0	

Ref: TEMP-0830 Rev.39 BACK / VERSO 1300060626

<b>LOT</b>		<b>MX 416 H</b>	<b>CONTROL H</b>							 (Exp.)	<b>2019-06-05</b> (YYYY-MM-DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>19.6</b>	<b>19.9</b>								± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.64</b>	<b>5.70</b>								± 0.20
HB HGB	g/dl	<b>17.9</b>	<b>18.1</b>								± 0.6
	g/l	<b>179</b>	<b>181</b>								± 6
	mmol/l	<b>11.12</b>	<b>11.24</b>								± 0.37
HT HCT	%	<b>49.6</b>	<b>50.2</b>								± 3.0
	l/l	<b>0.496</b>	<b>0.502</b>								± 0.030
VGM MCV	µm <sup>3</sup> ; fl	<b>88</b>	<b>88</b>								± 4
TGMH MCH	pg	<b>31.7</b>	<b>31.8</b>								± 2.0
	fmol	<b>1.97</b>	<b>1.97</b>								± 0.12
CCMH MCHC	g/dl	<b>36.1</b>	<b>36.1</b>								± 3.0
	g/l	<b>361</b>	<b>361</b>								± 30
	mmol/l	<b>22.40</b>	<b>22.41</b>								± 1.86
IDR RDW	%	N/A	<b>13.0</b>								± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>520</b>	<b>535</b>								± 45
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>8.5</b>								± 2.0
LYMPHO	#	N/A	<b>2.39</b>								± 1.00
	%	N/A	<b>12.0</b>								± 5.0
MONO	#	N/A	<b>0.90</b>								± 0.80
	%	N/A	<b>4.5</b>								± 4.0
GRANULO	#	N/A	<b>16.62</b>								± 1.40
	%	N/A	<b>83.5</b>								± 7.0